| Notice of Allowability | Application No. | Applicant(s) | |
|--|---|---------------------------------|---|
| | 09/898,936 | DICARLO, PAUL | |
| | Examiner | Art Unit | |
| | Christopher D. Prone | 3738 | |
| The MAILING DATE of this communication appears on the cover sheet with the correspondence address All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS. This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308. | | | |
| 1. This communication is responsive to the Appeal Brief Filed 3/14/05. | | | |
| 2. The allowed claim(s) is/are <u>1-15,25-27 and 35-52</u> . | | | |
| 3. The drawings filed on <u>03 July 2001</u> are accepted by the Examiner. | | | |
| 4. Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) | | | |
| Attachment(s) 1. Notice of References Cited (PTO-892) 2. Notice of Draftperson's Patent Drawing Review (PTO-948) 3. Information Disclosure Statements (PTO-1449 or PTO/SB/0 Paper No./Mail Date 4. Examiner's Comment Regarding Requirement for Deposit of Biological Material | 5. Notice of Informal Pa 6. Interview Summary Paper No./Mail Date 8), 7. Examiner's Amendm 8. Examiner's Stateme 9. Other | (PTO-413), e nent/Comment | · |

Reasons for Allowance

The following is an examiner's statement of reasons for allowance: The claimed combination of the prosthetic comprising a graft having hems on its ends and a cord disposed within the graft, that is adapted to expand upon absorbing fluid is considered patentable. The art of record provides for a graft with hems and stents, but they fail to disclose a cord that expands upon exposure to fluids. It is not deemed obvious to one of ordinary skill in the art to combine the art of record with a fluid expanding cord.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Christopher D. Prone whose telephone number is (571) 272-6085. The examiner can normally be reached on Monday Through Fri 8:30 to 5:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Corrine McDermott can be reached on (571) 272-4754. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Christopher D Prone Examiner Art Unit 3738

ύρ∙ CDP

> CORRINE McDERMOTT SUPERVISORY PATENT EXAMINER TECHNOLOGY CENTER 3700